



December 7, 2018

A review of our QC253, Full Range CO-Oximeter Control, lot# 84857 Exp Date: 2019-09, lot# 84953 Exp Date: 2019-10, and lot# 85049 Exp Date: 2019-10, has prompted us to adjust our assigned value for the AVOXimeter® 1000E, AVOXimeter® 4000 and GEM OPL analyzers. Please see the chart below for the corrected values.

QC253 Expected Values - AVOX1000e			
Level 1 Lot 84857	Mean	Min	Max
THb	8.1	7.4	8.8
O2Hb	93.5	89.0	98.0
Level 2 Lot 84953	Mean	Min	Max
THb	13.4	12.3	14.5
O2Hb	83.3	79.0	87.6
Level 3 Lot 85049	Mean	Min	Max
THb	16.1	14.8	17.4
O2Hb	55.0	50.7	59.3

QC253 Expected Values - AVOX4000 / GEM OPL			
Level 1 Lot 84857	Mean	Min	Max
THb	8.0	7.3	8.7
O2Hb	94.3	89.8	98.8
COHb	6.4	2.4	10.4
MetHb	0.9	-2.0	3.8
Level 2 Lot 84953	Mean	Min	Max
THb	13.5	12.4	14.6
O2Hb	83.6	79.3	87.9
COHb	17.2	12.7	21.7
MetHb	0.35	-2.6	3.3
Level 3 Lot 85049	Mean	Min	Max
THb	15.9	14.6	17.2
O2Hb	54.6	50.3	58.9
COHb	46.3	41.0	51.6
MetHb	0.2	-2.7	3.1

RNA Medical has not reprinted the insert for this lot, so please use this letter as notification for this change. Keep this letter with your records as long as you have this lot in stock. This letter will also be available on our website.

Please do not hesitate to contact RNA Medical at 1-800-533-6162 or by email at [CustomerService@RNAMedical.com](mailto:CustomerService@RNAMedical.com) if you have any questions regarding this revision.

BIOMKT-207  
REV. B