



May 20, 2021

Dear Valued Customer,

Instrumentation Laboratory would like to provide values for the Glucose and Lactate analytes for the following lot on the GEM Premier 3000 and 3500 series instruments.

QC823 Kit Lot: 115434 Exp: 2024-03-31

QC Level (Lot#)	Glucose		Lactate	
	Mean	Range	Mean	Range
1 (11534)	70	63-77	0.5	0.1-0.9
2 (11633)	182	164-200	2.3	1.9-2.7
3 (11734)	287	260-314	6.8	5.8-7.8

The insert that comes with the kit will not contain this information. This letter will serve as notification for the value assignment of these analytes for this lot. Please keep this announcement with your records for as long as you have this lot in stock.

If you have any questions regarding this notification, please feel free to call or email us using the contact information provided.

RNA Medical, Division of Bionostics, Inc.
7 Jackson Road
Devens, MA 01434
Phone: 978-772-9070 / 800-533-6162
E-Mail: customerservice@rnamedical.com

BIOMKT-224 Rev A