

INTENDED USE

RNA Medical® Brand **CVC 123** Calibration Verification Controls are assayed quality control materials used for confirming the calibration and linearity of blood gas, electrolyte, and metabolite instrumentation. This product may be used on stand-alone blood gas analyzers or any combination blood gas, electrolyte, and metabolite system. When used as a supplement to instrument calibration, daily QC procedures, preventative maintenance, and proper record keeping, CVC 123 will contribute to the laboratory's overall quality assurance program.

CVC 123 may be used to:

- Check analyzer linearity
- Troubleshoot analyzer problems
- Document proper calibration
- Confirm preventative maintenance
- Assist with regulatory compliance

PRODUCT DESCRIPTION

CVC 123 is a buffered aqueous solution containing electrolytes, glucose, and lactate. It has been equilibrated with specific levels of oxygen, carbon dioxide, and nitrogen. This product contains no preservatives and no biological materials.

CVC 123 is provided in convenient, ready-to-use ampuls containing 2.5 mL of solution. It is a five (5) level product with four (4) ampuls of each level in each kit. CVC 123 values cover the clinically significant range of instrument performance for pH, $p\text{CO}_2$, $p\text{O}_2$, Na^+ , K^+ , Cl^- , Ca^{++} , Mg^{++} , glucose, and lactate for documenting reportable range. For optimal performance, use of one (1) kit per analyzer is recommended.

STORAGE AND SHELF LIFE

CVC 123 has a shelf life of thirty-six (36) months from the date of manufacture when stored at 2-8 °C. It may be stored at room temperature (up to 25 °C) for nine (9) months, not exceeding the stated expiration date. This product should be protected from freezing and exposure to temperatures greater than 30 °C.

CONTROL VALUES AND ANALYTES

Lot specific values are provided with each box of controls. The typical target values for CVC 123 are as follows:

Analyte	Low Value		High Value
pH	6.85	-	7.80
$p\text{CO}_2$ (mmHg)	15	-	87
$p\text{O}_2$ (mmHg)	22	-	465
Na^+ (mmol/L)	90	-	170
K^+ (mmol/L)	1.6	-	11.3
Cl^- (mmol/L)	70	-	135
Ca^{++} (mmol/L)	0.25	-	3.10
Mg^{++} (mmol/L)	0.15	-	2.20
Glucose (mg/dL)	0	-	450
Lactate (mmol/L)	0.7	-	17.0

CVC 123 SPECIFICATIONS

Packaging: 5 levels (4 ampuls per level), 2.5 mL per ampul
 Storage: 2-8 °C (up to 9 months at room temperature)
 Shelf Life: 36 months from date of manufacture
 Matrix: Buffered aqueous solution
 Analytes: pH, $p\text{CO}_2$, $p\text{O}_2$, Na^+ , K^+ , Cl^- , Ca^{++} , Mg^{++} , glucose, lactate

DIRECTIONS FOR USE

The basic steps for running CVC 123 are outlined below. Refer to the package insert for specific instructions.

1. Calibrate the instrument as directed by the manufacturer.
2. Sample each of the Level 1 ampuls until three (3) replicates are completed. Test Levels 2, 3, 4, and 5 the same way.
3. Record the results on the Data Collection and Linearity Worksheets included in the kit.
4. Calculate the mean value for each test analyte and compare it to the range on the Expected Values Chart.
5. Using the graphs provided, plot the test result against the expected result. Connect the points to visualize linearity.

Note: Steps 3, 4, and 5 may be performed on-line as a feature of PeerQC® described below.

PEERQC® STATISTICAL ANALYSIS

Available at www.RNAMedical.com, PeerQC provides web-based graphing and reporting for RNA Medical's Calibration Verification Controls. This easy-to-use service eliminates time-consuming manual data calculation and hand linearity graphing.

Specific features of PeerQC include:

- Instant data calculation and analysis
- Peer comparison data in *real-time*
- Printable graph and detailed report for each analyte
- Report download for archiving

ORDERING INFORMATION

Please specify this catalog number when ordering RNA Medical CVC 123 Calibration Verification Controls.

Catalog number: CVC 123

ADDITIONAL PRODUCTS

RNA Medical offers a full line of blood gas and CO-Oximetry controls, calibration verification materials, capillary blood collection tubes, and quality control materials for various point-of-care tests. For further information about CVC 123 or any other RNA Medical product, please call us or visit our website at: www.RNAMedical.com.

RNA Medical is a registered trademark and PeerQC is a registered service mark of Bionostics, Inc., Devens, MA, USA.

INSTRUMENT REFERENCE CHART

CVC 123 Calibration Verification Controls

CVC 123

MANUFACTURER AND MODEL	ANALYTES REPORTED BY INSTRUMENT IN CVC 123
EPOCAL[®]	
EPOC [®]	pH, pCO ₂ , pO ₂ , Na ⁺ , K ⁺ , Ca ⁺⁺ , Glucose, Lactate
IL[®]	
1600 Series	pH, pCO ₂ , pO ₂ , Na ⁺ , K ⁺ , Ca ⁺⁺
Synthesis Series	pH, pCO ₂ , pO ₂ , Na ⁺ , K ⁺ , Cl ⁻ , Ca ⁺⁺ , Glucose
GEM [®] Premier [®] 3000 Series	pH, pCO ₂ , pO ₂ , Na ⁺ , K ⁺ , Ca ⁺⁺ , Glucose, Lactate
GEM Premier 4000	pH, pCO ₂ , pO ₂ , Na ⁺ , K ⁺ , Cl ⁻ , Ca ⁺⁺ , Glucose, Lactate
ITC[®]	
IRMA [®] and IRMA TRUpoint [®]	pH, pCO ₂ , pO ₂ , Na ⁺ , K ⁺ , Cl ⁻ , Ca ⁺⁺ , Glucose, Lactate
MEDICA	
EasyBloodGas	pH, pCO ₂ , pO ₂
EasyStat [®]	pH, pCO ₂ , pO ₂ , Na ⁺ , K ⁺ , Cl ⁻ , Ca ⁺⁺
NOVA	
CCX	pH, pCO ₂ , pO ₂ , Na ⁺ , K ⁺ , Cl ⁻ , Ca ⁺⁺ , Mg ⁺⁺ , Glucose, Lactate
pHOX [®]	pH, pCO ₂ , pO ₂
pHOX Plus	pH, pCO ₂ , pO ₂ , Na ⁺ , K ⁺ , Cl ⁻ , Ca ⁺⁺ , Mg ⁺⁺ , Glucose, Lactate
Stat Profile 1-10	pH, pCO ₂ , pO ₂ , Na ⁺ , K ⁺ , Cl ⁻ , Ca ⁺⁺ , Glucose, Lactate
Stat Profile Ultra	pH, pCO ₂ , pO ₂ , Na ⁺ , K ⁺ , Cl ⁻ , Ca ⁺⁺ , Mg ⁺⁺ , Glucose, Lactate
OPTIMEDICAL[®]	
OPTI [®] CCA	pH, pCO ₂ , pO ₂ , Na ⁺ , K ⁺ , Cl ⁻ , Ca ⁺⁺ , Glucose
OPTI R	pH, pCO ₂ , pO ₂ , Na ⁺ , K ⁺ , Ca ⁺⁺
OPTI LION	pH, Na ⁺ , K ⁺ , Cl ⁻ , Ca ⁺⁺
RADIOMETER[®]	
ABL [®] 5	pH, pCO ₂ , pO ₂
ABL 50, 500 Series	pH, pCO ₂ , pO ₂ , Na ⁺ , K ⁺ , Cl ⁻ , Ca ⁺⁺
ABL 600 Series	pH, pCO ₂ , pO ₂ , Na ⁺ , K ⁺ , Cl ⁻ , Ca ⁺⁺ , Glucose, Lactate
ABL 700 Series	pH, pCO ₂ , pO ₂ , Na ⁺ , K ⁺ , Cl ⁻ , Ca ⁺⁺ , Glucose, Lactate
ABL 77, 80 Series	pH, pCO ₂ , pO ₂ , Na ⁺ , K ⁺ , Cl ⁻ , Ca ⁺⁺ , Glucose
ABL 800 Series	pH, pCO ₂ , pO ₂ , Na ⁺ , K ⁺ , Cl ⁻ , Ca ⁺⁺ , Glucose, Lactate
NPT 7	pH, pCO ₂ , pO ₂
ROCHE[®]	
Compact Series	pH, pCO ₂ , pO ₂
Cobas [®] b 221	pH, pCO ₂ , pO ₂ , Na ⁺ , K ⁺ , Cl ⁻ , Ca ⁺⁺ , Glucose, Lactate
Roche OMNI [®] 1-9	pH, pCO ₂ , pO ₂ , Na ⁺ , K ⁺ , Cl ⁻ , Ca ⁺⁺ , Glucose, Lactate
SIEMENS[®] (BAYER[®])	
238	pH, pCO ₂ , pO ₂
248	pH, pCO ₂ , pO ₂
400 Series	pH, pCO ₂ , pO ₂ , Na ⁺ , K ⁺ , Cl ⁻ , Ca ⁺⁺ , Glucose
800 Series	pH, pCO ₂ , pO ₂ , Na ⁺ , K ⁺ , Cl ⁻ , Ca ⁺⁺ , Glucose, Lactate
600 Series	pH, Na ⁺ , K ⁺ , Cl ⁻ , Ca ⁺⁺
1200 Series	pH, pCO ₂ , pO ₂ , Na ⁺ , K ⁺ , Cl ⁻ , Ca ⁺⁺ , Glucose, Lactate
YSI[®]	
2300 Stat Plus	Glucose, Lactate

EPOCAL and EPOC are registered trademarks of Epocal Corporation, Ottawa, Ontario, Canada. Synthesis and Premier are trademarks and GEM is a registered trademark of Instrumentation Laboratory, Bedford, MA, USA. ITC, IRMA and IRMA TRUpoint are registered trademarks of International Technidyne Corporation, Edison, NJ, USA. EasyStat is a registered trademark of Medica Corporation, Bedford, MA, USA. Stat Profile and pHOx are registered trademarks of Nova Biomedical Corporation, Waltham, MA, USA. OPTI and OPTIMEDICAL are registered trademarks of Opti Medical Systems, Inc., Roswell, GA, USA. ABL is a registered trademark of Radiometer Medical APS Corporation, Bronshoj, Denmark. Cobas is a registered trademark of Roche Diagnostics Operations, Inc., Indianapolis, IN, USA. Roche OMNI is a registered trademark of Roche Diagnostics GMBH, Mannheim, Germany. Siemens is a registered trademark of Siemens Aktiengesellschaft FED REP GERMANY Wittelsbacherplatz 2 80333 München FED REP GERMANY. Bayer is a registered trademark of Bayer AG, Leverkusen, Germany. YSI is a registered trademark of YSI Incorporated, Yellow Springs, OH, USA.

BIOMKT-0006
Rev. D 8/2011